

Certificate of Conformity

Date: 03 May 2019

Product: Thermo Scientific Nunc Serological Pipette 25ml
Item Number: 170357
Lot Number: M330330
Date of Expiry: 04/2024

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
Accuracy	± 1% of nominal value	Complies
USP<85> Bacterial endotoxins test	<0.01 endotoxins unit/ml	Complies
USP <88> "Biological Reactivity test In Vivo, USP Class VI Test"	USP current edition	Complies**

**Only performed once

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10⁻⁶.

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:

J Mallett

Batch Release-Document Controller