

Certificate of Conformity

Date: 07 Oct 2019

Product: Thermo Scientific Nunc Serological Pipette 25ml
Item Number: 170357
Lot Number: M355130
Date of Expiry: 09/2024

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

| Test Description | Requirements | Result |
|--|----------------------------|------------|
| Visual | According to specification | Complies |
| Test of function | According to specification | Complies |
| Accuracy | ± 1% of nominal value | Complies |
| USP<85> Bacterial endotoxins test | <0.01 endotoxins unit/ml | Complies |
| USP <88> "Biological Reactivity test In Vivo, USP Class VI Test" | USP current edition | Complies** |

**Only performed once

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10⁻⁶.

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:

J Mallett

Batch Release-Document Controller